



EN Suction Unit

www.rossmax.com

Warranty Card

This instrument is covered by a 2 year guarantee from the purchase date. The guarantee is valid only on presentation of the guarantee card completed by the dealer confirming purchase date or the receipt. Suction unit accessories are not included. Opening or altering the instrument invalidates the guarantee. The guarantee does not cover damage, accidents or noncompliance with the instruction manual. Please contact your local seller/dealer or www.rossmax.com.

Customer Name: _____

Address: _____

Telephone: _____

E-mail address: _____

Product Information

Date of purchase: _____

Store where purchased: _____

Symbols

	Class II device
	Type BF applied part
	Attention: check the instructions before using this device
	More vacuum
	Device serial number
	Risk: Electrocution. Consequence: Death. Do not use the device while taking a bath or a shower
	Switch on
	Switch off
	Direct current
	CE Marking medical ref. Dir 93/42/EEC amended by 2007/47/EC
	Date of manufacture
	Manufacturer
	Attention: Consult accompanying documents
	WARNING: The symbol on this product means that it's an electronic product and following the European directive 2012/19/EU the electronic products have to be disposed on your local recycling centre for safe treatment.

Electromagnetic Compatibility Information

- This device needs to be installed and put into service in accordance with the information provided in the user manual.
- WARNING: Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the V7, including cables specified by the manufacturer. Otherwise, degradation of the performance of this device could result.

If higher IMMUNITY TEST LEVELS than those specified in Table 9 are used, the minimum separation distance may be lowered. Lower minimum separation distances shall be calculated using the equation specified in 8.10.

Manufacturer's declaration-electromagnetic immunity			
The V7 is intended for use in the electromagnetic environment specified below. The customer or the user of the V7 should assure that is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment-guidance
Conducted RF IEC 61000-4-6	3 Vrms: 0,15 MHz – 80 MHz 6 Vrms: in ISM and amateur radio bands between 0,15 MHz and 80 MHz 80 % AM at 1 kHz	3 Vrms: 0,15 MHz – 80 MHz 6 Vrms: in ISM and amateur radio bands between 0,15 MHz and 80 MHz 80 % AM at 1 kHz	Portable and mobile RF communications equipment should be used no closer to any part of the V7 including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance: d = 1,2 √P, d = 1,2 √P 80MHz to 800 MHz, d = 2,3 √P 800MHz to 2,7 GHz Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m). Interference may occur in the vicinity of equipment marked with the following symbol:
Radiated RF IEC 61000-4-3	10 V/m 80 MHz – 2,7 GHz 80 % AM at 1 kHz	10 V/m 80 MHz – 2,7 GHz 80 % AM at 1 kHz	

NOTE1: At 80 MHz and 800 MHz, the higher frequency range applies.
NOTE2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

More information on EMC compliance of the device can be obtained from Rossmax website: www.rossmax.com.

Microbial contamination

In the presence of pathologies with the risk of infection and microbial contamination, we recommend personal use of the accessories, collection canister, connecting tubing and antibacterial filter and all accessories should be thoroughly cleaned and disinfected after each use.

Note: Use only with original Rossmax accessories.

Accessories (optional)

Name	Model No.
Connecting PVC tubing (Ø7mm-length, 180cm)	SU_AC_001_00
Collection canister (1400ml)	SU_AC_002_00
Antibacterial filter kit (Antibacterial filter + 2 Silicone tubing Ø8mm-length, 14cm)	SU_AC_004_00
Antibacterial filter	SU_AC_005_00
Connecting Silicone tubing (Ø8mm-length, 14cm)	SU_AC_006_00
Tubing depository	SU_AC_008_00
Rechargeable Lithium-ion battery	SU_AC_009_00
AC power cord (EU plug)	SU_AD_004_00
AC power cord (US plug)	SU_AD_005_00
AC power cord (UK plug)	SU_AD_006_00
AC power cord (JET plug)	SU_AD_007_00
AC/DC switching Adaptor (12V/5A)	SU_AD_008_00 (PART No.: UES65-120500SPA1)

Technical Specifications

Suction class:	High Vacuum / High Flow
Adaptor Input:	100-240V, 50/60Hz; Output: DC 12V 5.0A
Max. Vacuum @ pump inlet:	-80 kPa/ -600 mmHg (mmHg for reference only)
Max air flow @ pump inlet:	30L/min (normal operation) and 25L/min (power saving/ economical mode)
Power Consumption:	<55W
Collection canister capacity:	1,400 ml
Battery charge duration:	approx. 45 mins under normal operation / approx. 60 mins under power saving/ economical mode at a zero vacuum level (free flow).
Battery charging time:	5 hours (at temp. 25°C)
Battery Rating:	7.2V 2900mAh*2 41.76WH
Weight:	Approx. 3.6 (Kgs)
Machine dimension:	363 (W) × 175 (D) × 254.5 (H) mm (Incl. rubber pad)
Operating conditions:	Temperature Min. 10°C; Max. 40°C; Relative humidity : 30% ~75%, non-condensing; Atmospheric pressure: Min. 700 hPa; Max. 1060 hPa
Storage and transportation conditions:	Temperature Min. -20°C; Max. 50°C; Relative humidity: 10~85%, non-condensing; Atmospheric pressure: Min. 700 hPa; Max. 1060 hPa
IP Classification:	IP21: Protection against harmful ingress of water and particulate matter

Troubleshooting

If, after checking the described conditions below the device is still not working properly, we recommend that you consult your local distributor or authorized Rossmax service center.

Problem	Possible cause	Solution
The device does not work	The AC adaptor has not been correctly inserted into the socket of the device or the power socket	Correctly insert the AC adaptor into the sockets
	Power switch has not been activated.	Check if the power switch is turned to the "I" (ON) position
The unit does not work from its battery	Battery flat	Fully charge the battery
	Battery is not installed or inserted properly	Reinstall battery and check if the battery connector is inserted properly
	Battery ages	If the unit does not operate after charging, replace the battery with new one
Lack of suction	Battery broken	If the battery cannot be fully charged and the GREEN charging indicator keeps flashing, replace the battery with new one.
	The collection canister lid has not been correctly positioned on the canister	Correctly position the collection canister lid
Blocked float	Lid gasket (rubber O-ring) not in place	Correctly position the gasket on the lid
	Leaks or cracks in canister assembly	Replace the canister
	Filter blocked	Replace the filter
Vacuum power poor and/or non-existent	Incrustation on the float	Remove the canister lid and overflow valve and extract the float. Continue by carrying out the cleaning procedures as described in the paragraph "Sanitization, Disinfection operations"
	Vacuum flow adjustment knob/ vacuum regulator completely open	Fully close the adjustment knob and check the vacuum power
	Filter blocked	Replace the filter
	Connecting tubing to the device or collection canister is clogged, bent, disconnected or broken; or suction catheter is twisted or blocked	Check the condition of the tubing, replace it if blocked or broken and correctly connect it as per the Assembling Procedures of this manual
	Canister lid overflow valve closed or blocked	Unblock the overflow valve, keep the device upright
Dirty, blocked or damaged pump	Take the device to your local distributor or authorized Rossmax service center	

In the event that contaminated fluids have moved past the internal unit, remove the tubing and the filter, clean the casing, and decontaminate the collection canister. Return the device to your local distributor or authorized Rossmax service center following **"Rules for returning and repairing"** paragraph.

1. Sanitization

Before and after each use, sanitize the collection canister and the connecting tubing.

Rinse each part (9,11,15,15a,16,16a,17,20,22) with clean, hot running water (<50°C) with gentle dishwashing liquid (non-abrasive). After cleaning, leave the parts to dry in an open, clean environment.

2. Disinfection

The accessories which can be disinfected are 9,11,15,15a,16,16a,17,20.

Wash with a commercial disinfectant that must be an electrolytic chloroxidizer (active principle: sodium hypochlorite) specific and follow the instructions and dilution rates supplied by the disinfectant manufacturer carefully. It is effective on the parts that undergo this treatment only if the parts to be treated have previously been sanitized. After cleaning, leave the parts to dry in an open, clean environment.

Parts	Material	Sanitization	Disinfection	Steam Sterilization (Gravity autoclave)
Collection Canister (9)	PC(Polycarbonate)	○	○	X
Canister supporting handle (17)	ABS / PC	○	○	X
Canister Lid (15)	ABS / PC	○	○	X
Gasket (15a)	Silicone	○	○	X
Overflow valve (16)	PP	○	○	X
Rubber valve (16a)	Silicone	○	○	X
Float (16a)	PP	○	○	X
Tubing depository (11)	PC	○	○	X
Connecting tubing (Ø7mm, 180cm) (22)	PVC	○	X	X
Connecting tubing (Ø8mm, 14cm) (20)	Silicone	○	○	X

○: Applicable, X: Not applicable

Note:

- The collection canister and tubing after disinfecting should be dated and stored in clean and dry areas to avoid recontamination.
- Do not use any cleaners or disinfectants that contain ammonia, benzene and/or acetone to clean the parts.
- Dispose of the solution following the instructions provided by the disinfectant manufacturer.
- Dispose of the suction catheter in accordance with the local laws and regulations.

Collection canister – 1400 ML

The collection canister consists of an overflow valve, a bottle in PC (polycarbonate) material, a lid with switch button on handle and must be replaced if there are visible signs of the damage.

Note:

- Keep the device vertical to allow the overflow valve to function correctly.
- Do not overturn the collection canister during use, in order to prevent the intervention of the overflow valve; should this occur, switch the aspirator off and detach the tubing connected to the antibacterial filter.
- To accelerate aspiration and simplify cleaning, it is recommended to put water in the collection canister before using.
- Suction at the patient is automatically obstructed when the liquid level reaches the overflow valve located on the underside of the canister lid.
- Contents of the container should be emptied when automatic overflow valve is activated. Further suctioning could cause damage to the vacuum pump.

Filter

The antibacterial filter supplied with Rossmax aspirators must be strictly replaced for each new patient, or if overflow occurs or the filter is saturated. Should the aspirator be used on the same patient, the filter for single-patient use is recommended to be replaced after every application. It should also be replaced if the unit is used with a patient whose risk of contamination is unknown. The filter cannot be sanitized, disinfected, or sterilized.

Note:

- Do not use the aspirator without the antibacterial filter, the unit could be damaged. From a bacteriological point of view, it becomes dangerous for the patient.
- Do not substitute any other material for this antibacterial filter. Substitution may lead to contamination or poor performance; use only Rossmax filters.

Silicone and PVC tubing

The tubing is recommended to be single use, but the number of cycles of disinfection and/or cleaning is strictly linked to the employment of the tubing. Therefore, after each cleaning cycle, it is up to the final user to verify whether the tubing is suitable for reuse. The component must be replaced if there are visible signs of the decay of the tubing.

Maintenance

The suction equipment does not need maintenance or lubrication. It is, however to inspect the unit and accessories before each use.

⚠ DANGER:

Electric shock hazard. Do not attempt to open or remove cabinet, there are no user-serviceable internal components. If service is required, return the device to your local distributor or authorized Rossmax service center. Opening or tampering with the unit will void warranty.

Rules for returning and repairing

Rossmax requires the following procedures to be carried out to protect the instrument and the safety of all who come into contact with it.

Before returning an instrument for repair, the external surfaces and all accessories **MUST** be carefully disinfected with a cloth soaked in methylated spirits or hypochlorite-based solution. The instrument and accessories should then be placed in a bag with a note outlining the disinfection undertaken. Failure to follow this procedure will result in the instrument being returned to the purchaser unrepaired. Instruments returned for repair **MUST** be accompanied by a description of the problem. Rossmax will not be responsible for damage caused through improper use.

To avoid such damage, please read the instruction carefully.

Rossmax will not be responsible for damage of accessories. Accessories will be replaced at the customer's expense.

Introduction

We thank you for purchasing our Aspirator V7 which is adaptable to your needs. It is very important that the health care worker and/or the patient read and understand the information for use and maintenance.

Intended Use

Suction unit (Aspirator) V7 can be used to remove unwanted fluids or infectious materials from the airway or respiratory support system in health care structures like hospitals and home care. The device generates suction (aspiration) which allows fluids to be extracted through connecting PVC tubing connected to a collection canister which retains the fluids until they can be properly disposed of. Use of the device must be prescribed by a doctor. Carefully follow the recommended operations and procedures outlined in this instruction guide to ensure proper functioning and maximize the life of the device.

Preliminary Remarks

This Suction unit (Aspirator) complies with the European regulations and bears the CE mark "CE 1639". The quality of the device has been verified and conforms to the provisions of the EC council directive 93/42/EEC amended by 2007/47/EC (Medical Device Directive), Annex I essential requirements and applied harmonized standards. EN ISO 10079-1:2015 Medical suction equipment - Part 1: Electrically powered suction equipment.

Warnings and Safety Instructions

Suction unit (Aspirator) V7 is EMC-tested in conformity with the requirements of IEC 60601-1-2 and can be used in the vicinity of other EMC-tested devices that fulfill the requirements as outlined in the IEC 60601-1-2 standard. Untested HF sources, radio networks or the like can impair the function of the device and should not be operated in combination with the Aspirator V7. The suction equipment is not suitable for use in an MRI environment. Please read and observe these warning and safety instructions before operation. These instructions for use must be kept with the device for later reference. Please note that these instructions are a general guide for the use of the product. Medical matters must be addressed by a physician. Rossmax considers himself only responsible for the effect on BASIC SAFETY, reliability and performance of the Aspirator V7 if it is used in accordance with the instructions for use.

⚠ Important Safeguards

This is a medical device and must be used by qualified users. It must be operated as indicated in this user instruction manual.

📖 Read All Instructions Before Using This Device. Save These Instructions.

The manufacturer makes every effort to ensure that every product is of the highest quality and safety; however, as for any electrical appliances, basic safety regulations must always be observed in order to avoid harming persons and things.

DANGER, WARNING, CAUTION, NOTE/ATTENTION STATEMENT


⚠ DANGER: Urgent safety information for hazards that will cause serious injury or death.

⚠ WARNING: Important safety information for hazards that draws attention to a potential danger.

⚠ CAUTION: Indicates correct operating or maintenance procedures to prevent damage to the product or other property.

⚠ Note/Attention: Indicates information to which users should pay special attention.

⚠ DANGER:

- The device is intended exclusively for the collection of NON-flammable or NON-corrosive fluid materials in medical applications only. This unit should not be used in the presence of flammable anaesthetic mixture with air, or with oxygen or nitrous oxide.
- Do not use outdoors or operate where aerosol (spray) products are being used or where oxygen is being administered in a closed environment such as an oxygen reservoir.
-  Do not handle the AC/DC switching adapter plug with wet hands or use the unit when taking a bath or shower. Never submerge the device in any liquids.
- Do not use the unit if the plug or AC adaptor is worn or wet and remove or touch the immersed device. If by chance it falls into water, do not attempt to remove the device from the water while the plug is still connected; pull the plug out of the socket and disconnect the mains switch. Bring the device to your local distributor or authorized Rossmax service center immediately. Do not attempt to make the device work before it has been thoroughly checked by qualified personnel and/or authorized Rossmax service center.
- Do not place or store the device where it can fall or be pulled into a tub or sink.

⚠ Warning:

- The medical device must NOT be used in the operating theater, for drainage or thoracic drainage and nasogastric suction.**
- Never operate this product if a) it has a damaged cord or plug, b) it is not working properly, c) it has been dropped or damaged, d) it has been dropped into water. Immediately return the product to your local distributor or authorized Rossmax service center for examination and repair.
- For use only by medically trained persons who have been adequately trained in suction procedures and in the use of aspirators.
- Incorrect use can cause pain and injury to the patient.
- Do not start the machine if you suspect the presence of explosive or inflammable liquids or gases, water inside the machine or environmental conditions that are not within the specified limits.
- Do not perform any maintenance operations while the device is being used on a patient.
- Replacement device must always be available if a breakdown of the device for patients can lead to a critical situation (e.g. patients with acute dyspnoea or severe catarrhal congestion),
- The patient should be monitored regularly according to the physicians instructions and facility guidelines. Objective indications or signs of a possible infection or complication must be met immediately (e.g. fever, pain, redness, increased warmth, swelling or purulent discharge). Non-observance can lead to considerable danger for the patient. Monitoring the Aspirator V7 for operating status is a must.
- Consult the indications for use and consider risk factors and contraindications before using the Aspirator V7. Failure to read and follow all instructions in this manual prior to use may result in serious or fatal injury of the patient.
- In the presence of children and non self-sufficient individuals, the device must be used under the close supervision of an adult who has been adequately trained in suction procedures and in the use of aspirators.
- Some parts of the device are small enough to be swallowed by children; keep the device out of the reach of children.
- Do not use the provided tubing and AC adaptor for any other purpose than those specified, as they can cause risk of strangulation. Be particularly attentive with children and individuals with disabilities because they are often unable to correctly assess risk.

- Repairs must be done only by your local distributor or authorized Rossmax service center. Unauthorized repairs void the warranty and may be hazardous for the user.
Attention: Do not modify this device without authorization from Rossmax.
- The filter used by the same patient is recommended to be replaced after every application.
Attention: The antibacterial filter must be replaced for each new patient.
- The suction catheter and the manual regulator are sterile, disposable products for single use and must be replaced after every application. Check the expiry date on the original packaging of the suction catheter and check the integrity of the sterile packaging. If expired and/or deteriorated, replace it at once.
Attention: The suction catheter and manual regulator are not supplied by Rossmax. Rossmax declines any responsibility for harm to the patient correlated to the deterioration of the aforementioned sterile packaging due to THIRD PARTY handling during the original packaging of the entire device
- Do not use sterile accessories when the sterile packaging is damaged.
- Do not connect this device to a passive drainage tube.
- Do not operate the device on an obstructed, unstable surface to avoid incidents, which may cause it to accidentally fall and lead to a malfunction and/or breakage. Should there be signs of damage to the plastic parts, which may expose inner parts of the device, do not connect the plug to the electrical socket. Do not attempt to make the device work before it has been thoroughly checked by qualified personnel and/or authorized Rossmax service center
- Never obstruct the air inlet positioned on the device.
- Use of the device in ambient conditions other than those specified in the manual may seriously impair its safety and technical characteristics.
- Before cleaning the device, pull the plug out of the fixed mains socket.
- The connecting tubing supplied with the canister must never come into direct contact with the suction area.
- Never use while drowsy or asleep.
- Always unplug the product immediately after use.
- Use the device only in dust-free conditions, otherwise treatment could be compromised.

⚠ Caution:

- Use only for the purpose intended.** Do not use for anything other than the use defined by the manufacturer. The manufacturer will not be responsible for damage due to improper use or connection to an electrical system not complying with current regulations.
- Do not tip the canister over while the device is working as the liquid may be aspirated into the device causing pump damage. Should this occur, immediately switch off the suction machine, then empty and clean the canister. Immediately bring the device to your local distributor or authorized Rossmax service center.
- Ensure that the connecting tubing and the canister lid have been carefully sealed in order to avoid leakage of suction.
- When the overflow valve intervenes, suction is halted. Switch off the device, empty the canister and carry out the cleaning procedures.
- In the event of aspiration without the canister and/or antibacterial filter, or if it is suspected that substances have entered the aspiration circuit, contact your local distributor or authorized Rossmax service center at once.
- Portable and mobile RF communications equipment can affect medical devices.
- The AC adaptor and rechargeable battery provided have been specifically designed for use with Rossmax series aspirators. Never use the AC adapter and rechargeable battery with other devices or for uses other than those specified in this manual, and never use Rossmax series aspirators with other power supply units.
- Before connecting the device always check that the electric data indicated on the data label and the type of plug used.
- The device casing is not waterproof. Do not wash the device under running water or by immersion and keep it safe from being sprayed by water or other liquid.
- Keep the AC/DC switching adaptor away from animals (for example, rodents) which could damage the insulation.
- Keep the device and AC/DC switching adaptor away from sources of heat, direct sunlight or extreme hot locations.
- When the liquid level reaches 50% of the total volume of the collection canister, it is recommended to empty or change it before the overflow valve acts, in this way the risk of the entrance of humidity in the antibacterial filter is minimized.
- The Rechargeable Lithium-ion battery integrated in the device is not to be considered as an ordinary domestic waste. Such a component must be disposed of in a specific collection center in order to be recycled.

Instructions for Use

The device must be checked before each use in order to detect malfunctions and / or damage caused by transport and / or storage and do not connect to power if damage is apparent. Before each use, the accessories must be carefully inspected to ensure the absence of dust, incrustation, clots or liquid substances both inside the connecting tubing and the collection canister. Furthermore, they must be cleaned following the instructions rigorously as stated in the "cleaning, sanitization, disinfection". We recommend personal use of the accessories, the collection canister, the connecting tubing and antibacterial filter to prevent risk of contagious infection.

1. Assembling procedures

- 1-1. Connect one end of connecting silicone tubing (Ø8mm,14cm), with the antibacterial filter marked **"IN"** side, to the VACUUM inlet of the collection canister marked in yellow
- 1-2. Connect the other end, with the antibacterial filter marked **"OUT"** side, to the "AIR INLET" connector of the suction unit marked in yellow.
- 1-3. Connect the connecting PVC tubing (Ø7mm, 180cm) to the "PATIENT" inlet of the collection canister.
- 1-4. Connect the power cord with AC adaptor, insert AC adaptor into the AC adaptor jack on the right side of the unit.

Note: Use only the authorized Rossmax AC adaptor with this aspirator.
Note: Ensure that the FLUID SIDE or **IN** marker on the filter is on the side facing the collection canister and fit into the **"VACUUM"** inlet. A wrong connection may cause immediate destruction in case of contact with sucked liquids.

2. Setup Before Use

- 2-1. Before each use, inspect the unit and accessories. If damaged, expired and/or deteriorated, replace it at once.
- 2-2. Place the unit on a flat, horizontal surface.
- 2-3. Plug the AC adaptor into the socket.
- 2-4. Place the switch to the "I" (ON) position.
- 2-5. Collapse the tubing on the antibacterial filter and check the vacuum gauge to test that the unit makes vacuum without any leakage and this level varies when the vacuum regulator is acted on.

2-6. Obstruct the connecting tubing and adjust the level of vacuum prescribed by the doctor in the vacuum gauge turning the vacuum flow adjustment knob/ vacuum regulator.

Note: The vacuum regulator can be used to set the level of vacuum required (mmHg/kPa). Turn the regulator clockwise, in the "+" direction, to obtain a higher vacuum; or anticlockwise for a lower vacuum; these values can be read on the vacuum gauge.

Important: The vacuum values on the control decal are purely for guidance; always refer to the reading shown on the vacuum gauge.

Note: Gauge is for reference only. If the unit sustains a severe drop, accuracy of the gauge must be checked.

2-7. Place the switch to the "O" (OFF) position.

Note: 1. The device is intended to be used on a flat, horizontal surface in order for the overflow valve to function correctly and prevent liquids from entering the suction pump. The device must be kept away from walls to allow for adequate engine cooling.
2. In case unit is dropped, always transport unit with vacuum regulator rotated fully clockwise.

3. Starting the Suction

3-1. Connect the suction catheter to the connecting PVC tubing (Ø7mm, 180cm).

Note: suction catheter and manual regulator are not supplied by Rossmax.

3-2. Switch on the unit (ON/OFF switch).

3-3. Place your finger on the manual regulator to activate the suction, begin suction on the patient using the suction catheter. The unit is ready to put the suction catheter to the patient.

3-4. After application, switch off the device and remove the AC adaptor from the mains socket. Perform the cleaning operations as described in the "cleaning, sanitization, disinfection " paragraph.

Note: A 30-minute interval is recommended after each use. Do not operate the device continuously for more than 30 minutes for a single use without turning it off and following with a cooling period for at least 30 minutes.

Power Source Options

• AC Operation: Connect the AC adaptor with the AC adaptor jack on the right side of the unit. Plug AC adaptor into a grounded wall-outlet power source.

• Battery Operation: The suction unit is equipped with a factory-installed and high-capacity rechargeable lithium-ion battery and has a LED light for low battery and charge indication. The device is supplied with the battery partially charged. We therefore recommend fully charging the battery for a minimum of 5 hours for initial use (please see battery charging).

LED






Switch the device on by pressing the "ON/OFF" button and the green battery indicator comes on. If the green light goes out during use, and the orange light comes on, there are about few minutes of battery charge left (reserve level). Therefore, terminate the application if possible. If the application cannot be terminated, it can be continued until the orange light flashes, but if this occurs do not use the appliance for more than 3 minutes to avoid damaging the battery. To continue the application with the battery flat, connect the unit to an external power source.

WARNING: If you get the low battery warning symbol, immediately switch to use an AC adaptor (external power source) to avoid an interrupted suction procedure. If the unit does not receive external power or the battery does not get recharged immediately, the low battery indicator keeps flashing and it will shut down rapidly.

Note:

- When the battery power drops below a set level the aspirator shuts down to avoid damaging the battery.
- To operate the unit from an internal rechargeable battery, ensure that no external power source (AC adaptor) is plugged into the suction unit. During charging or running of the unit, the power supply may become warm to the touch. This is normal.
- The AC adaptor can be used with the unit individually (without battery pack).

LED Explanations

	Battery charging: This GREEN light flashes while the battery is charging. The light will illuminate when the battery is fully charged.
	Low battery state: This ORANGE light flashes when the battery power is about to run out and must be charged as soon as possible.
	Middle battery state: This ORANGE light illuminates when there are about few minutes of battery charge left (reserve level).
	High battery state: This GREEN light illuminates when the battery has adequate power.
	Power Saving state: This GREEN light illuminates when the Powering Saving operation is activated.

Battery Charging

- Before using the rechargeable battery for the first time, charge the battery for a minimum of 5 hours to make sure it is fully charged.
- When charging the battery, use an external power source (AC adaptor) and verify that the Green charging indicator flashes when the power switch of the unit is in the **"O" (OFF)** position. The battery is **NOT** charged when the unit is in use, even the AC adaptor is plugged. Please be sure that the device you are using has a battery installed prior to contacting your local distributor if the battery is not charged.
- Discharging the battery completely will shorten the life of the battery. If the low battery ORANGE light flashes, do not operate the unit for more than 3 minutes. Recharge the battery as soon as possible.
- The GREEN charging indicator will keep flashing while the battery is charging.
- As the battery nears a full charge, the high battery state GREEN light will illuminate, not flash.

Note:

- Recharging the battery to full capacity may take up to 5 hours depending on the degree of discharge. The battery should be recharged every month if the unit is not in use for extended periods. The unit can be left charged when it is not in use to ensure that the battery is always charged to the optimal level. A fully-charged battery will provide approx. 45 minutes under normal operation and approx. 60 minutes under power saving operation/ economical mode. Operation time will decrease with higher vacuum levels.
- Unit run time will decrease as the battery ages.

3. Unit run time will also be reduced by letting the battery sit at a discharged state for extended periods.

Storage Note: Battery should be charged for a minimum of 5 hours prior to storage, and at least once every month.

Battery Replacement

⚠ Warning:


- If the fluid in the battery gets into your eye, wash the eye with sufficient water without rubbing the eye, then immediately consult the doctor for treatment. There is danger in losing your eyesight.
- Do not use the exclusive battery pack other than those provided with the unit, or it will void the warranty.
- Do not throw the battery pack into fire, or heat, or disassemble it. It may cause heat, ignition, short-circuit, or rupture.

⚠ Caution:

- Do not short-circuit the connector of the battery pack with a metal object such as wire.
 - If the fluid in the battery is stained on your skin or clothes, immediately wash off the fluid with water. You may suffer injury or the battery may leak, heat, ignite fire, or explode.
- Remove the screw on the battery cover located on the back side of the unit.
 - Disconnect the battery pack from the port and replace with a new one. Put the connector of battery pack into the port of battery bay.
 - Install the battery cover and fasten it with screw. At this time, be careful not to pinch the lead wire.
 - Connect the unit with the AC adaptor, and then charge the battery pack. The battery pack is partially charged. Before using the battery for the first time, charge the battery for a minimum of 5 hours to ensure it is fully charged.

Note:

- Battery replacement must be performed following the above procedures. If there are any problems, please contact your local distributor or authorized Rossmax service center for further help.
- It is recommended to replace the battery every 2 years, but the physical service life depends on the number of uses. If the unit does not operate for more than 10 minutes after fully charging, substitute the battery with a new one.
- Batteries or damage from old batteries are not covered by warranty.
- Batteries are hazardous waste. Do not dispose them together with the household garbage.



Power Saving Operation

Power saving operation can extend the operating time of rechargeable battery up to 60 minutes during service interruption and can also be used when low noise is needed.

- Switch on the unit (ON/OFF switch).
- Press the **"Power Saving"** key, and the indicator of Green light illuminates during use.
- To deactivate the power saving operation and return to normal operation, press the **"Power Saving"** key again, and the green light will go out.

Cleaning, Sanitization, Disinfection operations

Before using the device, the manufacturer advises you to clean and/or disinfect the accessories. Switch off the device before any cleaning procedures and unplug the AC adaptor from the socket.

Cleaning of the device

Use only a damp cloth with antibacterial soap (non-abrasive and with dissolvents of any sort) to clean the external surfaces of the main unit.

Caution:

- Particular care should be taken to ensure that the internal parts of the device do not get in touch with liquids.
- Do not submerge in water as this will result in damage to the vacuum pump.
- Do not use any cleaners or disinfectants that contain ammonia, benzene and/or acetone to clean the unit.

Collection canister and connecting tubing

- Detach the suction catheter (24), the manual regulator (23) and the connecting PVC tubing (22) from the collection canister.
- Disconnect the silicone tubing (20) from both the canister and the device.
- Remove the canister from its holder. Keep it upright, and empty it (in the WC at home, in the biological waste container in the hospital) and clean it after each use. Disassemble it as below:
 - Open the collection canister by pressing the switch button.
 - Remove the overflow valve (16) from the canister lid (15).
 - Remove the rubber valve and float (16a) from the overflow valve (16).
 - Remove the silicone gasket (15a) from the canister lid (15).

Sanitization and Disinfection operations

⚠ Warning:

- To prevent possible risk of infection from contaminated cleaning/disinfection solutions, always prepare fresh solution for each cleaning cycle and discard solution after each use.
- Contaminated devices must be thoroughly cleaned prior to disinfection.
- Personal protective equipment must be worn when handling contaminated equipment, particularly tubing, filter and collection canister (protective clothing, disposable latex gloves, mask, eye protection)
- If any liquid accidentally enters in the main unit or in the pump, the unit may be damaged. For this reason, we recommend not to use the aspirator. The failure should invalidate the guarantee under the manufacturer's responsibility.