



RM\_IB\_WE300  
EN\_SW\_ver203T  
IN00WE30000000000XX

Manual Instruction

Model: WE300

Manufacturer: Rossmax Swiss GmbH  
Widnauerstrasse 1, CH-9435 Heerbrugg, Switzerland  
www.rossmax.com

[www.cutie-rossmax.com](http://www.cutie-rossmax.com)

## Error Message

Lo

Low power : Please replace the battery immediately.

E r r

Overload : The weight of the baby is over 20 kgs.

## Technical data

1. Maximum capacity: 20kgs
  2. Minimum capacity: 100g
  3. Graduation: 5g
  4. Power source: 1x3V (CR2032) battery
  5. Operating environment: 0°C- 40; < 80% RH
- \* Subject to technical modifications.

## Warranty Card

This instrument is covered by a 2 year guarantee from the date of purchase, batteries and accessories are not included. The guarantee is valid only on presentation of the warranty card completed or stamped by the seller/dealer confirming date of purchase or the receipt. Opening or altering the instrument invalidates the guarantee. The guarantee does not cover damage, accidents or non-compliance with the instruction manual. Please contact your local seller/dealer or buying source or [www.rossmax.com](http://www.rossmax.com).

**Customer Name:** \_\_\_\_\_

**Address:** \_\_\_\_\_

**Telephone:** \_\_\_\_\_

**E-mail address:** \_\_\_\_\_

### Product Information

**Date of purchase:** \_\_\_\_\_

**Store where purchased:**

\_\_\_\_\_


## Introduction

Thank you for your decision to purchase Rossmax baby scale. Please peruse this instruction manual prior to first start up and carefully keep this user instruction so as to have this information on hand whenever it is needed.

## Safety guideline

- Always read the safety instructions carefully.
- Follow the basic safety precautions for all electronic equipment when using this scale.
- Keep this instruction manual for future reference.
- This unit is not waterproof. Don't expose this product to rain or moisture.
- Lay this equipment on a reliable flat surface before setting it up.  
If any of the following situation arises, get the equipment checked by the service technician:
  - The equipment has been exposed to moisture.
  - The equipment has been dropped and damaged.
  - The equipment shows signs of breakage.
  - Do not place the basic unit in a damp room at a distance of less than 1.5m away from a water source.
- Don't disassemble this scale. This scale doesn't contain any parts that can be repaired by the user.
- Use the scale only for its intended purpose.
- Clean the scale with a slightly damp cloth or with an anti-static cloth. Never use cleaning agents or abrasive solvents.

## Name/Function of each part



## Installing battery

1. Press down and lift the battery cover to open the battery compartment.
2. Install or replace one "CR2032" battery in the battery compartment according to the indications inside the compartment.
3. Replace the battery cover by clicking in the bottom hooks first, and then push in the top end of the battery cover.
4. Remove battery when unit is not in use for extended periods of time.

## You need to replace the battery when



1. low battery icon appears on display.
2. the ON/OFF/TARE key is pressed and nothing appears on display.

## Caution

1. Battery is hazardous waste . Do not dispose it together with the household garbage.
2. There are no user serviceable parts inside. Battery or damage from old battery is not covered by warranty.
3. Use exclusively brand battery. Always replace with new battery.

## Measurement procedure

1. Press the ON/OFF/TARE key to power on, the scale will appear zero "0.000" on the display after 2 seconds.
2. Put the baby on the weighing surface gently and carefully.
3. Under the measurement, the word "HOLD" is flashing on the display and stops flashing when measurement is completed.

## Automatic halt function

1. This scale has a "HALT" function for recording weight with struggling baby .
2. The automatic halt function can establish the baby's weight on the display as soon as he/she is still enough to achieve an exact measurement value.
3. As soon as the baby's weight is established on the display, the word "HOLD" stops flashing and the measured weight will be displayed for 20 seconds.
4. The scale will be switched off automatically.

## Manual halt function

1. If the baby is struggling or moving around vigorously, his/her weight cannot be recorded.
2. Press "HALT" button and the word "HOLD" will stop flashing within 3 seconds and the measured weight will be displayed for 20 seconds

## Tare function

1. If you use a towel to lay the baby on, it should be placed on the weighing surface before switching on.
2. If you forgot to do so, power on the scale, when it is ready to measure, place the towel on the weighing surface, and then press "ON/OFF/TARE" button to eliminate the weight of the towel.
3. Once you have pressed "ON/OFF/TARE" button, the word "TARE" will be shown on the display. (Note: If you remove the towel after press "ON/OFF/TARE" button, the scale will display a negative reading.)
4. To reset the scale to zero "0.000", press the "ON/OFF/TARE" button after the word "HOLD" has stopped flashing.

## Switching between kg/lb/oz units

Click the unit key in the bottom side of the scale to change the unit between "kg", "lb" and "oz" if necessary.